
Midmark M11 - v1.0.5
Total Cycles: 1907

Name: RITTER2
Sterilizer ID: V2313228
01/09/2023 02:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.3 F 0.0
2:00 177.6 F 3.3
4:00 193.0 F 8.0
6:00 212.4 F 13.5
8:00 222.7 F 8.4
10:00 234.1 F 11.2
12:00 240.8 F 12.2
14:00 241.1 F 11.4
16:00 255.0 F 18.6
18:00 267.1 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.3 F 29.5
2:00 271.5 F 29.4
3:00 272.2 F 29.6
4:00 271.3 F 29.0

Min 271.3 F 29.0
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 237.5 F 0.0
10:00 232.7 F 0.0
15:00 229.3 F 0.0
20:00 229.8 F 0.0
25:00 229.0 F 0.0
30:00 225.7 F 0.0

Drying Complete

FILLING: 1:39
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:09
CYCLE COMPLETE

01/09/2023 03:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____