

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.8uS / 0.5ppm  
CYCLE NUMBER 000847  
11:28 20/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:44  
PRESSURIZING 3:46  
STERILIZING 4:46  
136.0 C 220kPa 4:46  
135.9 C 221kPa 5:16  
136.4 C 222kPa 5:46  
136.4 C 224kPa 6:16  
136.4 C 224kPa 6:46  
136.4 C 224kPa 7:16  
136.4 C 224kPa 7:46  
136.4 C 224kPa 8:16  
136.4 C 224kPa 8:16  
VENTING 8:16  
AIR DRYING 9:13  
CYCLE COMPLETE 9:19

-----  
Digital Signature #  
3F75E17E2E53246A  
-----