

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001455  
9:23 18/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:02  
PRESSURIZING 2:11  
STERILIZING 3:20  
136.2 C 222kPa 3:20  
136.4 C 224kPa 3:50  
136.3 C 224kPa 4:20  
136.2 C 223kPa 4:50  
136.1 C 221kPa 5:20  
136.0 C 221kPa 5:50  
136.4 C 225kPa 6:20  
136.2 C 223kPa 6:50  
136.2 C 223kPa 6:50  
VENTING 6:50  
AIR DRYING 7:40  
CYCLE COMPLETE 22:40

-----  
Digital Signature #  
7D2290B311F4C1C8  
-----