
Midmark M11 - v1.0.5
Total Cycles: 2788

Name: RITTER2
Sterilizer ID: V2313228
08/18/2023 10:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.9 F 0.0
2:00 102.3 F 0.0
4:00 111.5 F 0.0
6:00 142.2 F 0.0
8:00 173.1 F 0.0
10:00 193.0 F 0.4
12:00 210.3 F 4.2
14:00 226.4 F 8.0
16:00 234.6 F 12.1
18:00 231.1 F 8.2
20:00 244.0 F 13.4
22:00 238.9 F 10.3
24:00 236.2 F 8.5
26:00 248.5 F 14.8
28:00 260.1 F 21.3
30:00 270.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.5 F 29.1
2:00 272.0 F 29.2
3:00 272.1 F 28.9
4:00 272.3 F 29.4

Min 271.4 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 237.3 F 0.0
10:00 191.8 F 0.0
15:00 181.9 F 0.0
20:00 178.8 F 0.0
25:00 175.9 F 0.0
30:00 174.8 F 0.0
Drying Complete

FILLING: 0:38
HEATING: 30:46
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:08:48
CYCLE COMPLETE

08/18/2023 11:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____