
Midmark M11 - v1.0.5
Total Cycles: 726

Name: RITTER2
Sterilizer ID: V2313228
01/25/2022 11:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.0 F 0.0
2:00 185.9 F 3.6
4:00 196.8 F 8.3
6:00 213.6 F 13.5
8:00 225.6 F 9.2
10:00 233.8 F 9.7
12:00 241.1 F 12.9
14:00 241.9 F 11.9
16:00 254.3 F 18.4
18:00 265.6 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.1 F 30.0
2:00 272.5 F 29.7
3:00 272.0 F 29.3
4:00 271.4 F 29.0

Min 271.2 F 28.9
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 210.9 F 0.0
10:00 211.6 F 0.0
15:00 217.0 F 0.0
20:00 217.9 F 0.0
25:00 224.0 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 19:52
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:17
CYCLE COMPLETE

01/25/2022 12:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____