
Midmark M11 - v1.0.0
Total Cycles: 5340

Name: RITTER 1
Sterilizer ID: V2178198
02/18/2025 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.5 F 0.0
2:00 68.2 F 0.3
4:00 82.1 F 1.6
6:00 117.0 F 3.8
8:00 145.9 F 6.3
10:00 173.0 F 10.1
12:00 201.8 F 5.9
14:00 209.5 F 4.8
16:00 224.5 F 8.2
18:00 234.3 F 11.9
20:00 229.8 F 7.0
22:00 241.3 F 12.1
24:00 238.2 F 9.5
26:00 251.3 F 15.8
28:00 262.6 F 22.3
30:00 272.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 271.7 F 28.3
2:00 272.0 F 28.6
3:00 272.2 F 28.8
4:00 272.4 F 28.9

Min 271.4 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 209.3 F 0.0
10:00 207.6 F 0.0
15:00 214.8 F 0.0
20:00 213.0 F 0.0
25:00 219.2 F 0.0
30:00 203.0 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:12
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:08:21
CYCLE COMPLETE

02/18/2025 02:40 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____