

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002295  
13:03 25/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:25  
PRESSURIZING 2:27  
STERILIZING 3:28  
136.0 C 220kPa 3:28  
136.1 C 221kPa 3:58  
136.1 C 219kPa 4:28  
CYCLE ABORTED 4:57  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 4:57  
-----

Digital Signature #  
B79EDD12546EFFFC  
-----