
Midmark M11 - v1.0.5
Total Cycles: 1707

Name: RITTER2
Sterilizer ID: V2313228
11/10/2022 11:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.2 F 0.0
2:00 150.5 F 0.1
4:00 170.7 F 0.7
6:00 198.0 F 3.2
8:00 217.9 F 7.8
10:00 230.0 F 12.7
12:00 232.5 F 9.0
14:00 240.5 F 10.5
16:00 243.6 F 12.9
18:00 243.5 F 12.2
20:00 256.2 F 19.1
22:00 268.1 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.8 F 29.7
2:00 271.8 F 29.0
3:00 272.3 F 29.8
4:00 272.6 F 29.5

Min 271.5 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 219.1 F 0.0
10:00 227.9 F 0.0
15:00 230.1 F 0.0
20:00 229.7 F 0.0
25:00 229.8 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:17
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

11/10/2022 12:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____