
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002483
15:12 06/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:32
STERILIZING 3:35
136.5 C 225kPa 3:35
136.3 C 224kPa 4:05
136.4 C 225kPa 4:35
136.4 C 225kPa 5:05
136.2 C 224kPa 5:35
136.4 C 226kPa 6:05
136.2 C 224kPa 6:35
136.3 C 223kPa 7:05
136.3 C 223kPa 7:05
VENTING 7:05
AIR DRYING 8:03
CYCLE COMPLETE 35:26

Digital Signature #
17A20362F9D360D5
