
Midmark M11 - v1.0.0
Total Cycles: 1777

Name: RITTER 1
Sterilizer ID: V2178198
03/31/2022 11:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.5 F 0.0
2:00 178.2 F 3.5
4:00 189.0 F 7.7
6:00 212.7 F 12.8
8:00 223.0 F 7.6
10:00 235.9 F 13.0
12:00 237.7 F 10.5
14:00 235.1 F 8.7
16:00 250.5 F 16.0
18:00 263.0 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.8
1:00 272.6 F 29.6
2:00 271.8 F 28.8
3:00 272.7 F 29.6
4:00 272.0 F 28.9

Min 271.5 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 214.0 F 0.0
10:00 223.3 F 0.0
15:00 231.8 F 0.0
20:00 232.2 F 0.0
25:00 230.6 F 0.0
30:00 227.7 F 0.0

Drying Complete

FILLING: 1:37
HEATING: 19:57
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

03/31/2022 12:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____