

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2666

Name: RITTER 1  
Sterilizer ID: V2178198  
12/13/2022 10:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.1 F 0.0  
2:00 190.0 F 2.8  
4:00 192.3 F 7.2  
6:00 212.1 F 12.9  
8:00 225.6 F 8.1  
10:00 241.4 F 12.3  
12:00 240.1 F 11.2  
14:00 238.0 F 9.4  
16:00 252.4 F 16.4  
18:00 264.6 F 23.5  
20:00 270.8 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.7 F 29.4  
2:00 272.5 F 28.9  
3:00 271.5 F 28.6  
4:00 272.7 F 29.4  
  
Min 271.5 F 28.5  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
5:00 242.9 F 0.0  
10:00 233.0 F 0.0  
15:00 223.9 F 0.0  
20:00 221.8 F 0.0  
25:00 226.9 F 0.0  
30:00 230.8 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 20:49  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:43  
CYCLE COMPLETE

12/13/2022 11:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_