
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 000294
7:58 13/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:14
STERILIZING 3:12
136.0 C 228kPa 3:12
136.4 C 230kPa 3:42
136.3 C 231kPa 4:12
136.4 C 231kPa 4:42
136.2 C 227kPa 5:12
136.2 C 231kPa 5:42
136.2 C 231kPa 6:12
136.2 C 230kPa 6:42
136.2 C 230kPa 6:42
VENTING 6:42
AIR DRYING 7:29
CYCLE COMPLETE 8:51

Digital Signature #
AA5E6245BCB2A86A
