
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001207
14:04 05/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:59
STERILIZING 4:08
136.0 C 220kPa 4:08
136.4 C 222kPa 4:38
136.3 C 222kPa 5:08
136.4 C 224kPa 5:38
136.2 C 222kPa 6:08
136.3 C 223kPa 6:38
136.4 C 224kPa 7:08
136.0 C 221kPa 7:38
136.0 C 221kPa 7:38
VENTING 7:38
AIR DRYING 8:29
CYCLE COMPLETE 23:29

Digital Signature #
B251408CEF5066C2
