
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005424
15:51 24/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:00
STERILIZING 3:51
136.4 C 223kPa 3:51
136.2 C 221kPa 4:21
136.4 C 224kPa 4:51
136.0 C 221kPa 5:21
136.4 C 223kPa 5:51
136.4 C 223kPa 6:21
136.1 C 221kPa 6:51
136.4 C 225kPa 7:21
136.4 C 225kPa 7:21
VENTING 7:21
AIR DRYING 8:12
CYCLE COMPLETE 11:06

Digital Signature #
6D9F5C8D88F52067
