
Midmark M11 - v1.0.5
Total Cycles: 2710

Name: RITTER2
Sterilizer ID: V2313228
07/27/2023 02:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.2 F 0.0
2:00 105.7 F 0.0
4:00 119.7 F 0.0
6:00 154.4 F 2.2
8:00 176.7 F 5.1
10:00 193.9 F 8.4
12:00 204.7 F 12.6
14:00 217.5 F 6.5
16:00 229.1 F 10.8
18:00 231.2 F 7.9
20:00 241.3 F 12.6
22:00 239.8 F 11.0
24:00 251.7 F 16.8
26:00 262.0 F 22.8
28:00 271.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.9 F 29.8
2:00 271.5 F 29.1
3:00 272.1 F 29.8
4:00 272.5 F 29.6

Min 271.2 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 205.2 F 0.0
10:00 190.8 F 0.0
15:00 187.6 F 0.0
20:00 185.8 F 0.0
25:00 182.9 F 0.0
30:00 180.8 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 28:38
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:06:58
CYCLE COMPLETE

07/27/2023 03:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____