
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000537
11:38 25/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:17
STERILIZING 4:31
136.0 C 220kPa 4:31
136.3 C 233kPa 5:01
136.4 C 233kPa 5:31
136.3 C 232kPa 6:01
136.4 C 230kPa 6:31
136.2 C 231kPa 7:01
136.3 C 232kPa 7:31
136.4 C 232kPa 8:01
136.4 C 232kPa 8:01
VENTING 8:01
AIR DRYING 8:55
CYCLE COMPLETE 23:55

Digital Signature #
12ECE851812709B2
