
Midmark M11 - v1.0.0
Total Cycles: 2541

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2022 01:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.5 F 0.0
2:00 191.3 F 4.7
4:00 194.9 F 9.7
6:00 215.7 F 6.2
8:00 229.5 F 9.7
10:00 235.0 F 8.3
12:00 243.1 F 12.9
14:00 242.5 F 11.6
16:00 255.6 F 18.5
18:00 267.5 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.2
2:00 271.7 F 28.5
3:00 271.6 F 28.9
4:00 272.3 F 29.1

Min 271.2 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 233.2 F 0.0
10:00 234.1 F 0.0
15:00 227.5 F 0.0
20:00 226.0 F 0.0
25:00 227.2 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:59
CYCLE COMPLETE

11/10/2022 02:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____