
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001522
10:51 10/21/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:30
PRESSURIZING 6:17
STERILIZING 7:48
136.0 C 220kPa 7:48
136.5 C 223kPa 8:18
136.2 C 221kPa 8:48
136.2 C 221kPa 9:18
136.4 C 222kPa 9:48
136.4 C 219kPa 10:18
136.2 C 221kPa 10:48
136.5 C 223kPa 11:18
VENTING 11:18
AIR DRYING 12:25
CYCLE COMPLETE 42:25

Digital Signature #
9961391E437D13FB
