
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007635
13:13 03/07/2024
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 4:35
STERILIZING 5:25
136.0 C 220kPa 5:25
136.4 C 225kPa 5:55
136.0 C 222kPa 6:25
136.3 C 224kPa 6:55
136.4 C 225kPa 7:25
136.4 C 224kPa 7:55
136.4 C 224kPa 8:25
136.1 C 223kPa 8:55
136.4 C 224kPa 9:25
136.3 C 223kPa 9:55
136.2 C 224kPa 10:25
136.3 C 224kPa 10:55
136.2 C 222kPa 11:25
136.1 C 222kPa 11:55
136.1 C 222kPa 12:25
136.4 C 226kPa 12:55
136.2 C 219kPa 13:25
136.3 C 224kPa 13:55
135.9 C 221kPa 14:25
136.4 C 222kPa 14:55
136.4 C 225kPa 15:25
136.4 C 225kPa 15:25
VENTING 15:25
AIR DRYING 16:17
CYCLE COMPLETE 35:29

Digital Signature #

E647F380FD652C03
