
Midmark M11 - v1.0.5
Total Cycles: 229

Name: RITTER2
Sterilizer ID: V2313228
07/21/2021 08:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.6 F 0.0
2:00 108.5 F 0.7
4:00 117.2 F 2.9
6:00 143.1 F 6.6
8:00 172.1 F 11.1
10:00 205.4 F 3.5
12:00 216.3 F 8.2
14:00 228.6 F 13.1
16:00 235.2 F 10.1
18:00 235.8 F 8.7
20:00 248.2 F 14.9
22:00 259.9 F 21.5
24:00 271.0 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.4
2:00 271.6 F 29.3
3:00 272.2 F 29.2
4:00 271.8 F 29.4

Min 271.3 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.0 F 0.0
5:00 218.4 F 0.0
10:00 213.8 F 0.0
15:00 199.3 F 0.0
20:00 195.1 F 0.0
25:00 191.7 F 0.0
30:00 190.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:44
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:02:45
CYCLE COMPLETE

07/21/2021 09:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____