
Midmark M11 - v1.0.5
Total Cycles: 1912

Name: RITTER2
Sterilizer ID: V2313228
01/10/2023 12:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.1 F 0.0
2:00 176.1 F 3.8
4:00 188.3 F 8.4
6:00 208.4 F 10.3
8:00 225.2 F 8.1
10:00 235.6 F 12.1
12:00 239.7 F 11.7
14:00 239.9 F 10.8
16:00 253.6 F 17.8
18:00 265.7 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.7 F 29.8
2:00 272.1 F 29.2
3:00 271.4 F 29.1
4:00 272.7 F 29.9

Min 271.3 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 238.3 F 0.0
10:00 233.5 F 0.0
15:00 231.5 F 0.0
20:00 227.3 F 0.0
25:00 230.4 F 0.0
30:00 227.5 F 0.0

Drying Complete

FILLING: 1:26
HEATING: 19:40
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

01/10/2023 01:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____