
Midmark M11 - v1.0.5
Total Cycles: 4200

Name: RITTER2
Sterilizer ID: V2313228
12/10/2024 03:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.6 F 0.0
2:00 91.2 F 0.0
4:00 117.8 F 0.2
6:00 151.6 F 3.0
8:00 175.0 F 7.1
10:00 200.7 F 11.4
12:00 213.7 F 4.7
14:00 226.1 F 8.9
16:00 237.6 F 13.5
18:00 238.3 F 10.7
20:00 235.9 F 8.9
22:00 250.0 F 15.6
24:00 261.8 F 22.4
26:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.6 F 29.2
2:00 272.3 F 29.4
3:00 272.0 F 28.8
4:00 272.8 F 29.6

Min 271.6 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 210.8 F 0.0
10:00 201.6 F 0.0
15:00 201.4 F 0.0
20:00 200.8 F 0.0
25:00 205.7 F 0.0
30:00 213.8 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:21
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:04:32
CYCLE COMPLETE

12/10/2024 04:37 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____