
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003595
15:32 03/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:14
STERILIZING 3:13
136.4 C 221kPa 3:13
136.2 C 223kPa 3:43
136.3 C 222kPa 4:13
136.4 C 224kPa 4:43
136.4 C 224kPa 5:13
136.4 C 225kPa 5:43
136.4 C 224kPa 6:13
136.3 C 223kPa 6:43
136.3 C 223kPa 6:43
VENTING 6:43
AIR DRYING 7:35
CYCLE COMPLETE 7:47

Digital Signature #
BDAE0E91321C18F5
