
Midmark M11 - v1.0.0
Total Cycles: 3524

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2023 11:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.0 F 0.0
2:00 196.7 F 3.3
4:00 202.8 F 7.9
6:00 226.0 F 13.5
8:00 230.2 F 8.6
10:00 239.6 F 11.3
12:00 241.6 F 11.8
14:00 240.2 F 10.6
16:00 254.0 F 17.5
18:00 265.9 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.5 F 29.0
2:00 271.4 F 28.6
3:00 272.5 F 29.1
4:00 271.4 F 28.4

Min 271.3 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 224.8 F 0.0
10:00 216.6 F 0.0
15:00 225.5 F 0.0
20:00 232.0 F 0.0
25:00 226.3 F 0.0
30:00 224.0 F 0.0

Drying Complete

FILLING: 1:48
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

07/13/2023 12:39 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____