

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4325

Name: RITTER 1  
Sterilizer ID: V2178198  
02/23/2024 08:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.0 F 0.0  
2:00 70.6 F 0.3  
4:00 77.2 F 1.1  
6:00 102.1 F 3.1  
8:00 133.3 F 5.7  
10:00 161.5 F 9.1  
12:00 182.6 F 12.7  
14:00 203.0 F 3.3  
16:00 212.5 F 6.2  
18:00 222.1 F 9.4  
20:00 230.6 F 13.2  
22:00 233.5 F 8.9  
24:00 242.3 F 12.1  
26:00 242.4 F 11.8  
28:00 253.9 F 17.6  
30:00 264.0 F 23.6  
32:00 271.9 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.5 F 29.0  
2:00 271.3 F 28.6  
3:00 272.2 F 29.1  
4:00 271.4 F 28.4  
  
Min 271.2 F 28.4  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.0 F 0.0  
5:00 220.2 F 0.0  
10:00 222.1 F 0.0  
15:00 237.2 F 0.0  
20:00 229.4 F 0.0  
25:00 226.5 F 0.0  
30:00 221.6 F 0.0  
  
Drying Complete

FILLING: 0:51  
HEATING: 32:10  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:10:31  
CYCLE COMPLETE

02/23/2024 10:03 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_