
Midmark M11 - v1.0.5
Total Cycles: 4309

Name: RITTER2
Sterilizer ID: V2313228
01/29/2025 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 71.1 F 0.0
4:00 79.5 F 0.1
6:00 113.8 F 1.3
8:00 138.2 F 3.9
10:00 161.7 F 7.1
12:00 183.5 F 10.4
14:00 203.5 F 8.1
16:00 217.3 F 5.6
18:00 228.1 F 9.0
20:00 237.2 F 12.9
22:00 234.5 F 9.4
24:00 238.7 F 10.5
26:00 244.4 F 13.2
28:00 255.5 F 19.1
30:00 265.8 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.3
1:00 272.6 F 30.1
2:00 271.8 F 29.2
3:00 272.7 F 29.9
4:00 271.7 F 29.2

Min 271.5 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 203.0 F 0.0
10:00 184.1 F 0.0
15:00 187.5 F 0.0
20:00 189.8 F 0.0
25:00 187.8 F 0.0
30:00 190.3 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:37
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:10:02
CYCLE COMPLETE

01/29/2025 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____