
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002453
8:55 27/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:45
STERILIZING 4:53
136.0 C 221kPa 4:53
136.4 C 225kPa 5:23
136.3 C 224kPa 5:53
136.3 C 223kPa 6:23
136.4 C 224kPa 6:53
135.9 C 221kPa 7:23
136.4 C 225kPa 7:53
136.1 C 221kPa 8:23
136.1 C 221kPa 8:23
VENTING 8:23
AIR DRYING 9:13
CYCLE COMPLETE 39:13

Digital Signature #
4736A46578B6C9A9
