
Midmark M11 - v1.0.0
Total Cycles: 2394

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2022 09:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.8 F 0.0
2:00 203.1 F 3.3
4:00 199.4 F 7.8
6:00 218.6 F 13.2
8:00 228.2 F 8.3
10:00 241.2 F 12.3
12:00 240.1 F 11.1
14:00 238.0 F 9.4
16:00 252.1 F 16.2
18:00 263.7 F 23.1
20:00 271.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.8
2:00 271.8 F 28.7
3:00 272.3 F 28.8
4:00 271.7 F 28.6

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 243.1 F 0.0
10:00 232.1 F 0.0
15:00 227.2 F 0.0
20:00 226.8 F 0.0
25:00 229.4 F 0.0
30:00 231.3 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:03
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

10/04/2022 10:34 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____