
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008184
11:41 19/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:41
STERILIZING 3:49
136.5 C 224kPa 3:49
136.4 C 224kPa 4:19
136.3 C 223kPa 4:49
135.8 C 220kPa 5:19
136.4 C 222kPa 5:49
136.4 C 223kPa 6:19
136.3 C 225kPa 6:49
136.3 C 223kPa 7:19
136.3 C 223kPa 7:19
VENTING 7:19
AIR DRYING 8:09
CYCLE COMPLETE 38:09

Digital Signature #
54F086C19D0B0C81
