
Midmark M11 - v1.0.0
Total Cycles: 1626

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2022 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.7 F 0.0
2:00 154.2 F 1.1
4:00 162.9 F 3.9
6:00 186.2 F 8.1
8:00 207.7 F 13.5
10:00 223.2 F 7.2
12:00 236.4 F 12.4
14:00 236.9 F 9.7
16:00 237.4 F 8.7
18:00 249.3 F 14.8
20:00 261.9 F 21.8
22:00 272.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.2
1:00 272.6 F 28.9
2:00 271.6 F 28.3
3:00 272.3 F 28.5
4:00 272.0 F 28.7

Min 271.5 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 207.7 F 0.0
10:00 206.4 F 0.0
15:00 205.6 F 0.0
20:00 203.9 F 0.0
25:00 204.5 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:00:13
CYCLE COMPLETE

02/14/2022 10:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____