
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003594
15:06 03/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:31
STERILIZING 3:29
136.4 C 223kPa 3:29
136.4 C 223kPa 3:59
136.3 C 224kPa 4:29
136.2 C 221kPa 4:59
136.2 C 223kPa 5:29
136.3 C 223kPa 5:59
136.4 C 224kPa 6:29
136.4 C 225kPa 6:59
136.4 C 225kPa 6:59
VENTING 6:59
AIR DRYING 7:50
CYCLE COMPLETE 8:02

Digital Signature #
3559B80E2465748D
