

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000052  
14:39 16/01/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:33  
PRESSURIZING 3:35  
STERILIZING 4:36  
136.0 C 220kPa 4:36  
136.4 C 221kPa 5:06  
136.3 C 225kPa 5:36  
136.0 C 222kPa 6:06  
136.2 C 224kPa 6:36  
136.3 C 224kPa 7:06  
136.4 C 225kPa 7:36  
136.3 C 224kPa 8:06  
136.3 C 224kPa 8:06  
VENTING 8:06  
AIR DRYING 9:00  
CYCLE COMPLETE 9:02

-----  
Digital Signature #  
958FB8FA8BFEA394  
-----