

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 007831  
13:14 06/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:35  
STERILIZING 3:37  
136.4 C 221kPa 3:37  
136.3 C 223kPa 4:07  
136.4 C 223kPa 4:37  
136.0 C 221kPa 5:07  
136.3 C 221kPa 5:37  
136.2 C 221kPa 6:07  
136.4 C 224kPa 6:37  
136.3 C 220kPa 7:07  
136.3 C 220kPa 7:07  
VENTING 7:07  
AIR DRYING 7:58  
CYCLE COMPLETE 20:08

-----  
Digital Signature #  
7B9C2995A6CE0DA6  
-----