

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4882

Name: RITTER 1  
Sterilizer ID: V2178198  
09/09/2024 05:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.3 F 0.0  
2:00 136.2 F 0.1  
4:00 155.7 F 0.4  
6:00 190.4 F 1.2  
8:00 210.7 F 5.3  
10:00 218.1 F 9.5  
12:00 232.1 F 10.2  
14:00 234.3 F 9.6  
16:00 237.4 F 9.5  
18:00 244.4 F 12.7  
20:00 242.5 F 11.5  
22:00 255.3 F 18.0  
24:00 266.4 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.7 F 29.0  
2:00 272.0 F 28.4  
3:00 271.7 F 28.5  
4:00 272.7 F 29.0

Min 271.5 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.4 F 0.0  
5:00 207.5 F 0.0  
10:00 210.2 F 0.0  
15:00 221.4 F 0.0  
20:00 229.9 F 0.0  
25:00 228.9 F 0.0  
30:00 233.3 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 25:40  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:04:11  
CYCLE COMPLETE

09/09/2024 06:23 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_