
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000487
11:48 28/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:23
STERILIZING 4:59
136.4 C 233kPa 4:59
136.4 C 232kPa 5:29
136.4 C 233kPa 5:59
136.4 C 232kPa 6:29
136.4 C 232kPa 6:59
136.3 C 232kPa 7:29
136.4 C 233kPa 7:59
136.2 C 230kPa 8:29
136.2 C 230kPa 8:29
VENTING 8:29
AIR DRYING 9:22
CYCLE COMPLETE 24:22

Digital Signature #
036ECD157C2D2FE4
