

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002298  
13:51 19/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:49  
PRESSURIZING 1:59  
STERILIZING 2:59  
136.4 C 223kPa 2:59  
136.0 C 222kPa 3:29  
136.4 C 224kPa 3:59  
136.3 C 225kPa 4:29  
136.3 C 224kPa 4:59  
136.3 C 223kPa 5:29  
136.4 C 226kPa 5:59  
136.4 C 226kPa 6:29  
136.4 C 226kPa 6:29  
VENTING 6:29  
AIR DRYING 7:20  
CYCLE COMPLETE 16:55  
-----

Digital Signature #  
1A93D4ADF98054D8  
-----