
Midmark M11 - v1.0.5
Total Cycles: 2887

Name: RITTER2
Sterilizer ID: V2313228
09/14/2023 04:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.8 F 0.0
2:00 173.4 F 3.1
4:00 188.1 F 7.3
6:00 205.4 F 11.9
8:00 218.2 F 6.4
10:00 229.8 F 11.4
12:00 233.0 F 9.0
14:00 238.2 F 10.4
16:00 246.1 F 14.1
18:00 257.9 F 20.5
20:00 268.7 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 271.7 F 29.9
2:00 272.5 F 30.0
3:00 271.6 F 29.5
4:00 271.2 F 29.4

Min 271.2 F 29.2
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.7 F 0.0
5:00 225.1 F 0.0
10:00 226.9 F 0.0
15:00 230.0 F 0.0
20:00 230.3 F 0.0
25:00 228.3 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:00:50
CYCLE COMPLETE

09/14/2023 05:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____