
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000466
12:41 02/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:54
STERILIZING 4:06
136.0 C 223kPa 4:06
136.2 C 231kPa 4:36
136.4 C 232kPa 5:06
136.2 C 232kPa 5:36
136.4 C 231kPa 6:06
136.3 C 232kPa 6:36
136.4 C 233kPa 7:06
136.4 C 233kPa 7:36
136.4 C 233kPa 7:36
VENTING 7:36
AIR DRYING 8:31
CYCLE COMPLETE 8:34

Digital Signature #
887C52654BB45461
