
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001516
15:17 10/20/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:55
PRESSURIZING 5:32
STERILIZING 6:55
136.0 C 220kPa 6:55
136.4 C 223kPa 7:25
135.8 C 218kPa 7:55
136.5 C 223kPa 8:25
136.0 C 221kPa 8:55
136.2 C 220kPa 9:25
136.4 C 224kPa 9:55
136.4 C 220kPa 10:25
VENTING 10:25
AIR DRYING 11:26
CYCLE COMPLETE 24:11

Digital Signature #
A1B2C55998001724
