
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
1.9uS / 1.2ppm
CYCLE NUMBER 000026
7:08 10/08/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 5:02
STERILIZING 6:11
136.0 C 223kPa 6:11
136.3 C 226kPa 6:41
135.8 C 221kPa 7:11
136.4 C 227kPa 7:41
136.5 C 228kPa 8:11
136.5 C 228kPa 8:41
136.4 C 226kPa 9:11
136.3 C 226kPa 9:41
VENTING 9:41
AIR DRYING 10:36
CYCLE COMPLETE 40:36

Digital Signature #
7D9E092507F1145D
