
Midmark M11 - v1.0.5
Total Cycles: 1890

Name: RITTER2
Sterilizer ID: V2313228
01/04/2023 08:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.3 F 0.0
2:00 76.8 F 0.0
4:00 81.4 F 0.0
6:00 106.4 F 0.0
8:00 144.0 F 0.1
10:00 176.4 F 1.5
12:00 196.0 F 4.8
14:00 209.8 F 8.6
16:00 221.3 F 12.5
18:00 228.9 F 7.5
20:00 237.9 F 12.1
22:00 236.6 F 9.7
24:00 238.1 F 9.4
26:00 247.7 F 14.5
28:00 259.3 F 21.0
30:00 269.9 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.3 F 29.3
2:00 272.0 F 29.7
3:00 271.7 F 28.9
4:00 272.4 F 29.7

Min 271.4 F 28.8
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.2 F 0.0
5:00 224.5 F 0.0
10:00 239.9 F 0.0
15:00 242.7 F 0.0
20:00 232.0 F 0.0
25:00 223.6 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:55
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:08:44
CYCLE COMPLETE

01/04/2023 09:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____