
Midmark M11 - v1.0.0
Total Cycles: 3052

Name: RITTER 1
Sterilizer ID: V2178198
03/29/2023 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.9 F 0.0
2:00 181.1 F 4.1
4:00 189.6 F 8.6
6:00 213.9 F 10.2
8:00 224.7 F 8.6
10:00 235.6 F 10.9
12:00 239.2 F 11.2
14:00 237.4 F 9.4
16:00 251.7 F 16.3
18:00 263.6 F 23.3
20:00 271.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.2
2:00 271.5 F 28.5
3:00 272.6 F 29.3
4:00 272.0 F 28.7

Min 271.4 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 222.0 F 0.0
10:00 238.3 F 0.0
15:00 236.9 F 0.0
20:00 223.5 F 0.0
25:00 220.4 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 20:03
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

03/29/2023 11:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____