
Midmark M11 - v1.0.5
Total Cycles: 3330

Name: RITTER2
Sterilizer ID: V2313228
01/29/2024 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.8 F 0.0
2:00 148.1 F 0.9
4:00 162.8 F 3.5
6:00 181.9 F 7.0
8:00 199.2 F 11.6
10:00 219.3 F 6.0
12:00 224.6 F 10.9
14:00 231.4 F 7.2
16:00 242.0 F 13.2
18:00 242.4 F 11.8
20:00 254.1 F 17.9
22:00 265.1 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.9
1:00 271.8 F 29.8
2:00 272.7 F 30.1
3:00 272.8 F 30.4
4:00 271.7 F 30.0

Min 270.4 F 28.7
Max 273.4 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 170.1 F 0.0
5:00 226.9 F 0.0
10:00 193.3 F 0.0
15:00 189.0 F 0.0
20:00 184.9 F 0.0
25:00 182.0 F 0.0
30:00 179.8 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:47
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:02:19
CYCLE COMPLETE

01/29/2024 12:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____