
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006955
12:50 16/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:54
STERILIZING 4:00
136.4 C 224kPa 4:00
135.9 C 222kPa 4:30
136.4 C 224kPa 5:00
136.2 C 223kPa 5:30
136.3 C 223kPa 6:00
136.0 C 222kPa 6:30
136.2 C 224kPa 7:00
136.4 C 223kPa 7:30
136.4 C 223kPa 7:30
VENTING 7:30
AIR DRYING 8:22
CASSETTE REMOVED
CYCLE COMPLETE 17:07

Digital Signature #
4C8E80E6FA2823F9
