

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1638

Name: RITTER2  
Sterilizer ID: V2313228  
10/21/2022 01:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.7 F 0.0  
2:00 172.7 F 3.5  
4:00 182.1 F 7.6  
6:00 205.4 F 12.8  
8:00 220.3 F 7.8  
10:00 233.7 F 13.1  
12:00 236.4 F 10.5  
14:00 235.6 F 9.0  
16:00 250.2 F 16.1  
18:00 262.6 F 23.2  
20:00 272.3 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.6 F 29.9  
2:00 271.8 F 29.3  
3:00 271.5 F 29.5  
4:00 272.7 F 30.0  
  
Min 271.2 F 29.1  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.6 F 0.0  
5:00 214.7 F 0.0  
10:00 233.2 F 0.0  
15:00 232.0 F 0.0  
20:00 226.8 F 0.0  
25:00 225.5 F 0.0  
30:00 214.9 F 0.0  
Drying Complete

FILLING: 1:25  
HEATING: 20:14  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:51  
CYCLE COMPLETE

10/21/2022 02:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_