

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3864

Name: RITTER 1  
Sterilizer ID: V2178198  
10/04/2023 04:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.3 F 0.0  
2:00 155.1 F 2.7  
4:00 175.3 F 6.2  
6:00 205.0 F 10.6  
8:00 216.9 F 4.3  
10:00 228.8 F 9.1  
12:00 236.3 F 10.4  
14:00 239.9 F 11.1  
16:00 237.4 F 9.3  
18:00 251.0 F 15.8  
20:00 262.5 F 22.4  
22:00 272.7 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.4 F 28.9  
2:00 272.8 F 29.3  
3:00 271.5 F 28.9  
4:00 272.4 F 29.0

Min 271.4 F 28.6  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.1 F 0.0  
5:00 192.4 F 0.0  
10:00 225.4 F 0.0  
15:00 228.4 F 0.0  
20:00 226.8 F 0.0  
25:00 226.8 F 0.0  
30:00 224.4 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 23:07  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:01:57  
CYCLE COMPLETE

10/04/2023 05:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_