
Midmark M11 - v1.0.5
Total Cycles: 2331

Name: RITTER2
Sterilizer ID: V2313228
05/03/2023 11:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.4 F 0.0
2:00 163.9 F 3.3
4:00 174.2 F 6.8
6:00 198.6 F 11.7
8:00 216.4 F 6.4
10:00 229.6 F 11.4
12:00 231.3 F 8.5
14:00 239.3 F 11.0
16:00 244.9 F 13.6
18:00 257.3 F 20.2
20:00 268.3 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.3 F 29.9
2:00 271.2 F 29.6
3:00 272.3 F 29.9
4:00 271.1 F 29.3

Min 271.1 F 29.3
Max 272.5 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 208.0 F 0.0
10:00 210.3 F 0.0
15:00 225.8 F 0.0
20:00 230.1 F 0.0
25:00 215.6 F 0.0
30:00 214.6 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 21:13
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:59:48
CYCLE COMPLETE

05/03/2023 12:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____