

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1692

Name: RITTER2  
Sterilizer ID: V2313228  
11/07/2022 05:51 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.4 F 0.0  
2:00 174.3 F 0.4  
4:00 195.8 F 3.8  
6:00 212.2 F 8.9  
8:00 226.5 F 9.5  
10:00 235.7 F 11.0  
12:00 235.9 F 9.3  
14:00 237.9 F 9.3  
16:00 251.0 F 16.1  
18:00 263.7 F 23.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.4 F 29.2  
2:00 271.6 F 29.0  
3:00 272.4 F 29.2  
4:00 271.4 F 28.7  
  
Min 271.4 F 28.6  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.1 F 0.0  
5:00 228.8 F 0.0  
10:00 230.7 F 0.0  
15:00 229.4 F 0.0  
20:00 227.2 F 0.0  
25:00 227.7 F 0.0  
30:00 225.8 F 0.0  
Drying Complete

FILLING: 1:30  
HEATING: 19:57  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:35  
CYCLE COMPLETE

11/07/2022 06:49 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_