

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4419

Name: RITTER2  
Sterilizer ID: V2313228  
03/11/2025 09:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.9 F 0.0  
2:00 76.3 F 0.3  
4:00 87.6 F 1.4  
6:00 120.9 F 4.0  
8:00 158.0 F 7.3  
10:00 181.5 F 10.9  
12:00 207.6 F 3.8  
14:00 218.7 F 6.5  
16:00 229.5 F 10.2  
18:00 231.0 F 9.0  
20:00 236.4 F 10.4  
22:00 235.3 F 8.6  
24:00 246.2 F 14.0  
26:00 257.7 F 20.1  
28:00 268.2 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 272.8 F 29.7  
2:00 271.7 F 29.3  
3:00 272.8 F 29.6  
4:00 271.6 F 29.3

Min 271.5 F 29.0  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
5:00 220.4 F 0.0  
10:00 231.1 F 0.0  
15:00 186.3 F 0.0  
20:00 186.4 F 0.0  
25:00 188.0 F 0.0  
30:00 191.1 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 29:18  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:07:45  
CYCLE COMPLETE

03/11/2025 10:30 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_