
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000159
12:04 15/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:09
STERILIZING 4:21
136.4 C 220kPa 4:21
136.2 C 224kPa 4:51
136.2 C 224kPa 5:21
136.3 C 225kPa 5:51
136.4 C 226kPa 6:21
136.4 C 226kPa 6:51
136.2 C 224kPa 7:21
136.4 C 225kPa 7:51
136.4 C 225kPa 7:51
VENTING 7:51
AIR DRYING 8:45
CYCLE COMPLETE 8:51

Digital Signature #
850660CF9E16EAEF
